



MAJESTIC



IQ-8000

Owner's Manual / Warranty

**COMFYGO**  
mobility

Version 2.1



# QUICK START GUIDE

- A. Open the box, pull out the chair, and conduct a visual inspection.
- B. Confirm that the controller cord is connected and secured to the battery.
- C. With the chair in a vertical position, insert the controller.
- D. Lay the chair flat with all four wheels on the ground and the controller facing upwards.
- E. Unfold your chair.
- F. Make sure both levers on the motor are in the "locked" position.
- G. Your electric wheelchair is ready to use.



- Don't forget to check under your chair for:
- Charger
  - Remote control
  - Owner's manual

Make sure both levers on the motor are in the "locked" position.



**! WARNING:** Please read this user manual carefully and check all of the safety tips.





# TABLE OF CONTENTS

1. Preface .....	03
2. Comfygo Care+ Protection .....	04
3. Tech Support .....	05
4. Claim Your Free Accessory .....	06
5. Technical Specs .....	07
6. Safety Instructions .....	08
7. Your Electric Wheelchair .....	13
8. Controller (Joystick) .....	14
9. Remote Controller .....	16
10. Usage & Operation .....	17
11. Folding & Unfolding .....	18
12. Footrest Adjustments .....	19
13. Headrest Installation .....	20
14. Battery & Charger .....	21
15. Features .....	23
16. Seat Extension Kit Installation .....	25
17. System Diagnostics .....	27
18. Maintenance Areas .....	30
19. Warranty Policy .....	32

- For all your spare parts needs, please visit **www.ComfyGoMobility.com**, email us at **info@comfygomobility.com**, or call **+1 844 755 5055**.
- Don't forget to **register your product to activate your manufacturer warranty**. Detailed information can be found on page 5.
- Please retain the box and all packaging for at least 30 days from the date of delivery.

**IMPORTANT NOTICE! BEFORE THE FIRST USE:**

**Ensure the battery is connected and charge it for at least 6 hours.**

To access additional information, simply scan the QR code with your smartphone's camera.



Register & Activate  
Your Warranty




Download the User Manual  
Watch Installation / Assembly Videos



# PREFACE

- This manual provides instructions on operation, assembly methods, and basic troubleshooting. It applies to our model: IQ-8000.
- This manual includes wheelchair maintenance tips and troubleshooting guidelines. Please keep it in an accessible location and refer to it as needed. If someone else uses the chair, please provide them with this manual for reference.
- The annotations and illustrations in this manual might differ slightly from your actual model as we continually enhance and upgrade our components and designs.
- If you're unsure about any aspect of the operation, kindly reach out to your dealer for clarification.
- Improper usage of this electric wheelchair might result in injuries. It's essential to always heed safety recommendations to ensure your well-being and that of others.
- This electric wheelchair is designed for individuals with walking difficulties or those who cannot walk.
- The electric wheelchair is intended to transport only one person.
- This series of electric wheelchairs are powered by Li-ion batteries and uses DC motors. Users can control the direction and adjust the speed using the joystick controls located on the arm. This wheelchair is suitable for low speeds, good road conditions, and can handle gentle slopes.

 **WARNING! Ensure the chair is locked when unfolded. Prior to the first use, please inspect all nuts and bolts, and tighten as necessary.**

- This electric wheelchair is specifically engineered for people weighing less than the maximum loading weight defined in the technical specs. It is intended for use at walking speeds and is suitable for both indoor and outdoor use. Please refrain from using this unit in traffic lanes.
- This owner's manual is compiled from the latest specifications and product information available at the time of publication. We reserve the right to make changes as they become necessary. Any changes to our products may lead to slight variations between the illustrations and explanations in this manual and the product you have purchased. The latest/current version of this manual is available on our website.



## Warning symbol

Follow the instructions next to this symbol closely.

Please pay attention to these instructions; otherwise, it may result in physical harm or damage to the user, the product, or the environment.





# COMFYGO CARE+ PROTECTION

- Your product comes with a limited manufacturer's warranty: 3 years for the frame, 1 year for the battery, and 1 year for the motor and electronics. More details can be found in the Warranty section on the following pages.
- Keep your mobility products protected and ensure a hassle-free experience with our ComfyGO Care+ Protection Plans.
- Designed for your peace of mind, these plans provide extended coverage for your ComfyGO mobility products, allowing you to focus on what matters most – *your independence and mobility*.

## Joystick (Controller)

Starting from \$249  
without coverage

## Engine

Starting from \$149  
without coverage

## Remote Controller

Starting from \$63  
without coverage



## Battery

Starting from \$249  
without coverage

## Wheel

Starting from \$89  
without coverage



## Get your protection plan today!

- Scan the QR code with your smartphone
- Call +1 844 755 5055
- Email us at [info@comfygomobility.com](mailto:info@comfygomobility.com)
- Visit our website at [ComfyGOMobility.com](https://ComfyGOMobility.com)

*Prices are for informational purposes only. For the most updated pricing, please visit our website, call, or email us.*





# TECH SUPPORT



### Register & Activate Your Warranty

- Please visit [Comfygomobility.com/pages/product-registration](https://comfygomobility.com/pages/product-registration) and register your product now. Simply scan the QR code with your phone to access the Registration Link.
- When you register your ComfyGO Product, you ensure that you may be eligible for extended warranty service.
- You will also receive product updates and important news from ComfyGO Mobility.

### Getting Tech Support

- For the quickest support, please email [info@ComfyGoMobility.com](mailto:info@ComfyGoMobility.com)
- Call us at **+1 844 755 5055**
- Visit [comfygomobility.com/pages/contact](https://comfygomobility.com/pages/contact) or scan the QR code with your smartphone to fill out the help form.



### Getting Tech Support in EUROPE

- For the quickest support, please email [info@ComfyGoMobility.eu](mailto:info@ComfyGoMobility.eu)
- Call us at **+31 (0) 20 809 7031**
- Visit [comfygomobility.eu/pages/contact](https://comfygomobility.eu/pages/contact) or scan the QR code with your smartphone to fill out the help form.

## Shield your wheelchair with the ComfyGO Travel Protection Bag

- Ensure the safety of your wheelchair during transit with our ComfyGO Travel Protection Bag. Tailored specifically for your product, this bag offers robust protection against scratches and damage while you travel.
- Be it in the hands of airline personnel or during the rigors of transportation, our Travel Protection Bag is your wheelchair's safeguard against the rough and tumble of travel handling.



### Get Your Travel Bag Today

- Call us at **+1 844 755 5055**
- Visit [comfygomobility.com](https://comfygomobility.com)
- Email us at [info@ComfyGoMobility.com](mailto:info@ComfyGoMobility.com)






# CLAIM YOUR FREE ACCESSORY

We're giving away a **FREE ComfyGO accessory** to enhance your ComfyGO Mobility experience!



1. Visit [comfygomobility.com](https://comfygomobility.com) and locate your product, or simply scan the QR code to access the page.
2. Write a review for your product.
3. Email [info@comfygomobility.com](mailto:info@comfygomobility.com) to inform our team about your review.
4. Our team will then provide you with accessory options.



Phone Holder



Headlight & Charger



Cup Holder



Side Bag

**Please also don't forget to leave a review at your point of purchase.**

## Terms and Conditions

- **Eligibility:** This promotion is open to individuals who have previously purchased a ComfyGO Mobility product, excluding accessories, parts, services, and protection plans. Employees of ComfyGO Mobility, its affiliates, subsidiaries, advertising, promotion, and fulfillment agencies, and legal advisors, as well as their immediate family members and persons living in their same household, are not eligible to participate in this promotion. Limit one entry per household.
- **Promotion Period:** This promotion is open for the first 1,000 customers who complete the review process and notify our team via email at [info@comfygomobility.com](mailto:info@comfygomobility.com). The promotion may end at the discretion of ComfyGO Mobility once a sufficient number of eligible entries have been received.
- **Promotion Modification:** ComfyGO Mobility reserves the right, at its sole discretion, to modify, suspend, or terminate this promotion at any time and for any reason without prior notice. Changes to the promotion, including but not limited to its termination date, will be updated on the ComfyGO Mobility website. Participants agree that ComfyGO Mobility shall not be liable to them or to any third party for any modification, suspension, or discontinuance of the promotion.
- **Selection of Winners:** Participants who have completed the steps mentioned above will be eligible for the giveaway. Our team will contact eligible participants to provide them with options for the free ComfyGO accessory.
- **Prize:** The prize consists of one ComfyGO accessory with a value equal to or less than \$25. The exact accessory will be determined based on available stock. No cash or other substitution of prizes is permitted, except at the sole option of ComfyGO Mobility for a prize of equal or greater value.
- **Shipping:** Winners are responsible for paying any shipping costs associated with the delivery of the prize. ComfyGO Mobility will not cover these costs.
- **Restrictions:** Reviews left on third-party marketplaces such as Amazon, Walmart, and others are not eligible for this promotion due to their review policies.
- **General Conditions:** Participants agree to abide by the terms of these Official Rules and by the decisions of ComfyGO Mobility, which are final on all matters pertaining to this promotion. Noncompliance with the foregoing, failure to abide by these Official Rules, or return of the prize notification as non-deliverable may result in disqualification and selection of an alternate winner.
- **Disclaimer:** ComfyGO Mobility, its agents, and representatives, are not responsible for and shall not be liable for: (i) late, lost, damaged, misdirected, incomplete, illegible, or unintelligible entries; (ii) any condition caused by events beyond the control of ComfyGO Mobility that may cause the promotion to be disrupted or corrupted; (iii) any injuries, losses, or damages of any kind arising in connection with or as a result of the prize, or acceptance, possession, or use of the prize, or from participation in the promotion; or (iv) any printing or typographical errors in any materials associated with the promotion.
- **Data Protection:** By participating in this promotion, participants agree to ComfyGO Mobility's use of their personal information for both online and offline direct marketing purposes.



# IQ-8000 TECHNICAL SPECS

Parameters	IQ-8000	IQ-8000 PLUS (20" Seat)
Overall Dimensions (L x W x H)	37"x23.5"x37" (939x597x940 mm)	38.5"x26.5"x39" (978x675x990 mm)
Folded Dimensions (L x W x H)	23.5"x15"x30" (597x380x762 mm)	26"x16"x30" (660x405x762 mm)
Seat Height (from ground)	20" (508mm)	20" (508mm)
Armrest Height (from seat)	9" (230mm)	9" (230mm)
Backrest Height (from seat)	20" (508mm)	20" (508mm)
Max Loading Weight	330 lbs (150 kg)	330 lbs (150 kg)
Net Weight without Battery	52.9 lbs (24 kg)	55 lbs (25 kg)
Net Weight with Battery	58.5 lbs (26.5 kg)	61 lbs (27.5 kg)
Battery	24V 12AH or 24V 20AH lithium-ion	24V 12AH or 24V 20AH lithium-ion
Battery Charge	AC 100-240V, 50Hz, 1.5-5A	AC 100-240V, 50Hz, 1.5-5A
Battery Charge Time	6 to 8 hours	6 to 8 hours
Range with 12AH Battery	Up to 13 mile (20km)*	Up to 13 mile (20km)*
Range with 20AH Battery	Up to 19 mile (30km)*	Up to 19 mile (30km)*
Airline and Cruise Approval*	Only with 24V 12AH battery**	Only with 24V 12AH battery**
Max Speed	4 mph (6 km/h)	4 mph (6 km/h)
Motor	2 * 250W (Total 500W)	2 * 250W (Total 500W)
Controller	24V 35A	24V 35A
Climbing Angle Max Degrees	6°	6°
Max Safe Slope	9°	9°
Static Stability	9°	9°
Brake	Intelligent electromagnetic break	Intelligent electromagnetic break
Turning Radius	≤ 47" (1200 mm)	≤ 47" (1200 mm)
Reversing Width	≤ 49" (1250 mm)	≤ 49" (1250 mm)
Front Wheel Diameter	8" (200 mm)	8" (200 mm)
Rear Wheel Diameter	12" (305mm)	12" (305mm)
Seating Area Width	17.5" (445 mm) ***	20.5" (520 mm) ***
Cushion Size (width x depth)	16" (410 mm) x 16" (410 mm)	18" (460 mm) x 17" (432 mm)
Controller Type	All-Direction Joystick	All-Direction Joystick
Tire pressure (for pneumatic tire)	30 psi (2 kgf/cm2)	30 psi (2 kgf/cm2)

\* The actual range of our product can vary due to factors such as weather conditions, load weight, and terrain type. Extreme temperatures, heavy loads, and challenging terrains can all impact battery efficiency and reduce travel distance. While designed for versatility, users should be mindful of these variables to optimize their range and maintain reliable performance.

\*\* The 12AH (288WH) lithium-ion battery is travel approved by most major airlines and cruise lines. Before travelling by airline or cruise, please contact the airline or cruise line to confirm their travel battery regulations.

\*\*\* The widest distance between armrest support bars.





# SAFETY INSTRUCTIONS

**IMPORTANT:** Before using your mobility device, please read and understand these safety instructions thoroughly to ensure your safety and the longevity of the device.

## **GENERAL SAFETY:**

- **User Limitations:** This device is intended for use by individuals with mobility impairments. Consult your healthcare provider to determine if this device is suitable for your condition.
- **Weight Capacity:** Do not exceed the maximum weight capacity as defined in the technical specifications section of this manual, as doing so may compromise the structure and safety of the device.
- **Inspection:** Regularly inspect your mobility device for signs of wear, loose components, or damage. If any issues are found, cease use immediately and contact the manufacturer or a certified technician for repair.
- **Modifications:** Do not modify the mobility device in any way, as this may affect its safety and void the warranty.

## **MAIN SECURITY FEATURES:**

- **Protection Against Electric Shock:** Internal power classified as Type B application.
- **Ingress Protection Rating:** Classified based on the degree of liquid ingress protection: IPx3
- **Flammability:** Classified as Non-AP/APG type for safe use in environments where flammable anesthetic mixtures with air or oxygen or nitrous oxide gases occur.
- **Operation Mode:** Designed for continuous operation.
- **Voltage Rating:** Operates at a DC 24V rating.
- **Defibrillator Discharge Protection:** Device has no protective effect on the application of defibrillator discharge.
- **Signal Ports:** No signal output or input part.
- **Non-Permanent Equipment:** Safety ensured for any non-permanent equipment installed.

## **ENVIRONMENTAL SAFETY:**

- **Terrain:** Use the device on stable and even surfaces. Avoid high water, icy, or uneven terrain that may cause the device to slip or tip.
- **Weather Conditions:** Avoid using the device in extreme weather conditions that may impair its function or your control over it.
- **Obstacles:** Be cautious of obstacles in your path and maintain a safe distance from edges, steps, and bodies of water.

## **RIDING NOTICE:**

### **General Riding Cautions:**

- Keep hands on the armrest for full control of the wheelchair.
- Practice in safe, open areas before using on roads.
- Assistive guidance is recommended until comfortable with chair control.
- Adhere to pedestrian traffic rules; the wheelchair is not an automobile.
- Stay on sidewalks and crosswalks; avoid vehicle lanes.
- Maintain steady speeds; avoid sharp turns and zigzag movements.
- Ensure proper tire pressure for safe use and battery longevity.





# SAFETY INSTRUCTIONS

- If your wheelchair supports reclining, DO NOT RECLINE the backrest while on any incline or decline.

## **Avoid Riding in These Conditions Unless Accompanied by a Caretaker:**

- Inclement weather (rain, fog, wind, snow, etc.)
- Wet wheelchairs should be dried immediately to prevent electronic damage.
- Challenging terrain (mud, dirt, sand, gravel, etc.)
- Crowded or busy roads.
- Railways (ensure wheelchair tires do not get stuck; turn off and cross carefully).

## **Precautions for Uphill and Downhill:**

- Avoid steep inclines and declines; the slope range should be less than that defined in the technical specifications section of this manual, as exceeding this may compromise the structure and safety of the device.
- Operate controller carefully during uphill use.
- Maintain forward momentum uphill or downhill; do not stop abruptly.
- Use brakes to control speed downhill.
- Avoid sideways movements on steep inclines.
- Do not use stairs; avoid crossing high steps.
- Cross ditches at a 90° angle.
- Keep wheelchair in powered mode during inclines and declines.
- Seek immediate assistance if wheelchair malfunctions at a railroad crossing; switch to manual mode and push wheelchair to safety.

## **Precautions for Caregivers**

- Ensure the user's feet are properly positioned on the footrests.
- Avoid clothes getting caught in the wheels.
- Push the wheelchair manually when on steep slopes.

## **OTHER IMPORTANT NOTICES:**

### **Repair and Refit:**

- Contact the seller or service department for repairs or refits.
- Do not modify the wheelchair; this could lead to accidents or malfunctions.
- Avoid adding non-approved weights or accessories.

### **Keep Your Wheelchair Dry at All Times:**

- Avoid storing or using the wheelchair in wet conditions.
- Dry the wheelchair immediately if it gets wet.

### **Pushing Someone Else in the Wheelchair:**

- Use only manual mode when pushing someone in the wheelchair.

### **Service Life:**

- The recommended service life of this product is 5 years from the date of production.
- Check the label on your unit box for the date of manufacture or contact us
- Avoid using beyond the recommended period to prevent accidents.





# SAFETY INSTRUCTIONS

## **ELECTROMAGNETIC COMPATIBILITY:**

- Avoid strong magnetic fields and large inductive electrical equipment.
- Maintain distance from sources of electromagnetic interference.
- Note that portable and mobile RF communication devices can affect wheelchair performance.
- Refer to manufacturer's statement for detailed electromagnetic compatibility information.
- Battery cut-off voltage: 23V.
- Classification: D class in GB/T 18029.21.

## **EMERGENCY PROTOCOL:**

Malfunction: If the device malfunctions, cease use immediately. If safe to do so, turn off the device and contact the manufacturer or a certified technician.

Accidents: In case of an accident, prioritize your safety and seek immediate assistance. Once safe, assess the device for damages.

## **CONTACT INFORMATION:**

- For further assistance or in case of emergency, please contact our customer support at:
- Phone: +1 844 755 5055
- Email: [info@comfygomobility.com](mailto:info@comfygomobility.com)
- Website: [www.comfygomobility.com](http://www.comfygomobility.com)
- Remember, adhering to these instructions ensures your safety and optimal performance of your mobility device. For hair battery, please contact your authorized provider.

**IMPORTANT:** Please read and follow all of these safety instructions and warnings to prevent injury or damage. Keep this notice for future reference.

## **General Usage Instructions and Warnings:**

- Understanding Your Power Chair: Before operating your power chair, ensure that you have read and understood the user manual, safety instructions, and guidelines provided.
- Weight Capacity Compliance: Do not exceed the specified weight capacity of your power chair. Overloading may result in instability, structural damage, or malfunction.
- Operating Environment: Operate your power chair on stable and even surfaces. Avoid steep slopes, near dropbar/kPa rating on each tire. Avoid underinflation or overinflation.
- Tire Care: Do not use high-pressure hoses or rubber conditioners on tires, as these can cause damage or reduce grip, leading to skidding.

## **Battery and Charging Instructions and Warnings:**

- Charging Guidelines: Follow the specific instructions for battery charging as outlined in the user





# SAFETY INSTRUCTIONS

manual and the information supplied with the battery charger.

- **Ventilation:** Charge your battery in a well-ventilated area. Explosive gases may be produced during charging.
- **Charger Compatibility:** Use only the off-board charger supplied with your power chair. Other types of chargers may cause damage.
- **Charger and Wiring Inspection:** Regularly inspect the battery charger, wiring, and connectors. Contact your authorized provider if you notice any damage.
- **Charger Handling:** Do not attempt to open or repair the battery charger. If malfunctioning, consult an authorized provider.
- **Charger Heat:** Be aware that the battery charger may become hot during charging. Avoid direct skin contact.
- **Charger Protection:** Protect the battery charger from adverse weather conditions if not rated for outdoor use. Allow it to adjust to indoor environmental conditions before use.
- **Battery Type:** Lithium-Ion batteries are recommended to mitigate risks of leakage or explosive conditions.
- **Charging Completion:** Monitor the LED lights on the charger for the status. If charging is incomplete within 24 hours, unplug and contact your provider.
- **Disposal:** Dispose of the power chair according to local and national regulations. Consult local waste disposal agencies or an authorized provider.

## **Maintenance and Storage Instructions and Warnings:**

- **Routine Checks:** Perform daily, weekly, and monthly checks as specified in the user manual. Pay special attention to tires, brakes, electrical connections, seating systems and that all nuts and bolts are tightly fastened.
- **Storage Conditions:** Store your power chair in a dry place, away from temperature extremes. Disconnect the battery and switch it to the "0" position during storage.
- **Storage Precautions:** Improper storage can lead to frame corrosion and electronic damage. Charge the battery periodically during prolonged storage.

## **Cleaning and Disinfection Instructions and Warnings:**

- **Cleaning Products:** Use a damp cloth and mild, non-abrasive cleanser for cleaning. Avoid products that may scratch the surface.
- **Disinfectant Use:** If using disinfectants, ensure they are safe for your product and follow all safety instructions to avoid skin irritation or damage to the power chair.

## **Service and Replacement Instructions and Warnings:**

- **Authorized Service:** Wheels, batteries, and other components should be serviced or replaced only by an authorized provider or a qualified technician.
- **Lifting Precautions:** Exercise caution when lifting heavy components like batteries and use proper lifting techniques.
- **Electrical Safety:** Keep tools and metal objects away from battery terminals to avoid electrical shock.





# SAFETY INSTRUCTIONS

## Electric Wheelchair Safety Usage Checklist

### Before Use:

- Ensure hands are positioned on armrests for optimal control.
- Verify tire pressure matches manufacturer recommendations.
- Check battery level.
- Inspect wheelchair for any signs of damage or wear.
- Confirm the chair is locked when unfolded, and all nuts are tightened.
- Make sure anti-tip wheels are properly assembled.

### During Use:

- Adhere strictly to pedestrian traffic regulations.
- Stay on sidewalks; avoid vehicle lanes at all times.
- Maintain steady, controlled speeds; abrupt maneuvers are prohibited.
- Avoid inclines and declines over 9 degrees.
- Keep feet firmly placed within the designated foot area.

### In Specific Conditions:

- If weather is adverse, ensure a caretaker is present.
- In case of water exposure, immediately dry the wheelchair.
- Seek assistance when navigating challenging terrains.
- Employ manual operation at railway crossings and verify tire clearance.
- Ascend and descend slopes with utmost caution, avoiding lateral movements.
- Avoid stair navigation and crossing over high obstacles.
- Ensure user's feet are properly placed on footrests.
- Keep user's clothing clear from the wheels.
- Provide assistance on slopes and in high-risk conditions.

### Post-Use:

- Store the wheelchair in a dry, safe environment.
- Conduct a brief inspection for potential damage or issues.
- Charge the battery if necessary, avoiding overcharging.

### Repair and Maintenance:

- Contact certified technicians for repairs or modifications.
- Report any malfunction or irregularities immediately.
- Schedule regular maintenance checks as recommended by the manufacturer.



*Note: This checklist is to be followed rigorously each time the electric wheelchair is used to ensure the safety of both the user and the equipment. Non-compliance with any of these points may lead to unsafe conditions and potential injury.*

*Liability Disclaimer: Adherence to these safety and operation guidelines is a strict requirement for all users and caregivers. Non-compliance or improper use of the wheelchair absolves the manufacturer of any liability for resultant injuries or damages. These regulations are implemented for the protection and well-being of all users and must be followed rigorously.*





# YOUR ELECTRIC WHEELCHAIR



- 1. Backrest
- 2. Controller (Joystick)
- 3. Controller Charging Port
- 4. Armrest
- 5. Cushion (Seat Base)
- 6. Storage
- 7. Rear Wheel (Drive Wheel)
- 8. Front Wheel ( Caster Wheel)
- 9. Foot Platform
- 10. Folding and Locking Mechanism
- 11. Battery Box
- 12. Motor
- 13. Anti-tip Wheel
- 14. Manual Freewheel Lever
- 15. Battery Charging Port





# CONTROLLER (JOYSTICK)



**Power Button:** To turn on the chair, press the power button. To turn it off, press the button again.

- When the chair is powered on, the LED Battery Power Indicator will light up.

**Sleep Mode:** If the joystick is not used for more than 20 minutes, the power will automatically turn off, and the system will enter sleep mode. The system can be turn on by pressing the power button.

**Horn Button:** Press the horn button to sound the horn.

**Mute Button:** Press the mute button once to silence the audio; press it again to restore sound.

**SOS Button:** In case of emergency, press and hold the SOS button to activate an audible alert. The SOS button will emit a continuous beeping sound until deactivated. Press the button again to turn off the alert.

**Speed Adjustment:** The wheelchair's riding speed can be adjusted according to the user's preference and circumstances.

- Use the decrease and increase buttons to adjust the speed accordingly.
- Speed is divided into five segments, corresponding to the following ranges: 20%, 40%, 60%, 80%, and 100% of the maximum speed.
  - The first LED represents 20% of the maximum speed.
  - The fifth LED represents 100% of the maximum speed.

**! WARNING!** In an emergency, you can also turn off the power directly by pressing the power button.





# CONTROLLER (JOYSTICK)

## LED Battery Power Indicator:

- The LED battery indicator is active when the power is on. It also displays the remaining battery power.
- If only the red or yellow LED lights are on, this signals low power and that the batteries need re-charging.
- Always ensure the batteries are fully charged before going on long rides where charging might not be possible.
- Do not operate the unit if only the red light is on, as it indicates the need for charging.
- LED 5: 80~100% battery remaining
- LED 4: 60~80% battery remaining
- LED 3: 40~60% battery remaining
- LED 2: 20~60% battery remaining
- LED 1: <20% battery remaining
- When the 1st LED blinks: <5% battery remaining.

## Using the Joystick

- To control the speed and direction of the wheelchair, utilize the joystick.
- The distance you push the joystick determines the speed of movement.
- Always ensure the joystick is in the middle position when powering the unit on or off. If the joystick is not centered when turning the unit on or off, an alert will sound. To silence the alert, return the joystick to the center. If the alert continues even after centering the joystick, this could signal a malfunction. In such a case, please contact your dealer.





# REMOTE CONTROLLER

**Please refer to the picture of the remote controller on page 15 of this user manual.**

## **Pairing the Remote with the Chair Controller**

1. If the chair is charging, unplug the charging cable.
2. Lay the chair flat with the wheels on the ground.
3. Power on the chair controller, then power on the remote by holding the power button for 3 seconds until it blinks red.
4. On the chair controller, press and hold the speed decrease button until the LED speed indicator displays a single amber light in the center of the screen.
5. Release the speed decrease button and press the horn.
6. On the remote, press and hold the speed decrease button until the chair controller turns off.
7. Power off the remote.
8. Power on both the remote and chair controller – the remote is now paired.

*For a video explanation on remote pairing, please visit the user manual page on our website. For detailed information, refer to the USAGE & OPERATION SECTION on the following pages.*

**Power Button:** To turn on the remote controller, press the power button. To turn it off, press the button again.

**Sleep Mode:** If the remote controller is not used for more than 20 minutes, the power will automatically turn off, and the remote controller will enter sleep mode.

**Horn Button:** Press the horn button to sound the horn.

**Speed Adjustment with Remote Controller:** The wheelchair's riding speed can be adjusted according to the user's preference and circumstances.

- Use the decrease and increase buttons to adjust the speed accordingly.
- Speed is divided into five segments, corresponding to the following ranges: 20%, 40%, 60%, 80%, and 100% of the maximum speed.
  - The first LED represents 20% of the maximum speed.
  - The fifth LED represents 100% of the maximum speed.

## **Using the Remote Controller's Joystick**

- To control the speed and direction of the wheelchair, utilize the joystick.
- The distance you push the joystick determines the speed of movement.
- Always ensure the joystick is in the middle position when powering the unit on or off. If the joystick is not centered when turning the unit on or off, an alert will sound. To silence the alert, return the joystick to the center. If the alert continues even after centering the joystick, this could signal a malfunction. In such a case, please contact your dealer.

**! WARNING:** *\*To use your remote control, the chair's joystick must be turned on first. The power button on the remote control only turns the remote control on and off. The remote control connects to the chair's joystick via Bluetooth.*






# USAGE & OPERATION



- Please visit our website for a detailed explanation on Usage and Operations.
- Go to [www.comfygomobility.com/pages/user-manuals](http://www.comfygomobility.com/pages/user-manuals) or scan the QR code with your smartphone's camera.
- Once on the website, select your wheelchair model and the video of interest.
- On the same page, you can also download the most updated digital version of your owner's manual.

### Initial Setup:

- **Unboxing:**
  - Carefully open the box.
  - Gently remove the chair from the box.
- **Visual Inspection:**
  - Conduct a thorough visual inspection of the chair to check for any visible damage or missing parts.



Charging  
Port

Battery  
Knob

### Controller Cord Connection:

- Confirm that the controller cord is properly connected and secured to the battery (as illustrated in the picture to the left).
- This is crucial for the functioning of the chair.

Controller Plug



Screw

### Attaching the Controller to the Armrest:

- Slide the controller into the armrest tube.
- Ensure that the controller is kept in a horizontal position (as illustrated in the to the left).
- Once the controller is properly positioned, secure it by tightening the screw (refer to the picture guidance).





# FOLDING & UNFOLDING

## Manual Folding Chair Instructions


If you have purchased a manual folding chair, please follow the instructions below on how to fold and unfold your chair.



Backrest Crossbar



Seat



Folding Lock

## Unfolding Your Wheelchair:

1. Grasp the backrest with one hand.
2. With your other hand, grab the seat to help unfold the chair.
3. Gently but firmly unfold the wheelchair until it is fully open.
4. Secure the lock located underneath the backrest. This step is crucial; ensure the lock is well fastened to prevent the wheelchair from folding unintentionally while in use.

## Folding Your Wheelchair:

1. Release the lock under the backrest. Make sure it's completely disengaged.
2. Hold the backrest with one hand.
3. With your other hand, pull the seat upwards to initiate the folding process.
4. Fold the wheelchair gently but firmly until it is completely closed.

*Note: Remember to always handle your wheelchair with care to ensure its longevity and your safety. If you encounter any difficulties or if anything seems amiss with the folding mechanism, it's best to consult the manufacturer or a professional for assistance.*




**WARNING:** Failure to lock the chair when unfolded is a safety violation.



# FOOTREST ADJUSTMENTS

## Footrest (Foot Platform) Adjustments

Your wheelchair comes with an adjustable footrest.



### Adjusting the Height of the Footrest:

- The footrest offers three height options.
- To adjust, first release the nuts and bolts to disconnect the footrest from the main frame.




- There are three adjustment holes available for different heights.
- Position the footrest at the desired height on the frame, ensuring that the holes are aligned. Then, reinsert all the nuts, bolts, and spacers in the same order as they were removed.
- Make sure the footrest is securely reconnected to the frame to prevent any safety issues.



### Adjusting the Extension of the Footrest:

- The footrest offers three extension options.
- To adjust, first release the quick-release nuts to disconnect the footplate from the footrest.



- There are three extension adjustment holes available for your comfort.
- Position the footplate at the desired extension on the footrest, ensuring that the holes are aligned.
- Then, reinsert both of the quick-release nuts.
- Make sure the footplate is securely reconnected to the footrest to prevent any safety issues.



# HEADREST INSTALLATIONS

Your wheelchair comes with an optional headrest available for purchase as an add-on. **Get Your Headrest Today** call us at **+1 844 755 5055** or visit **[comfygomobility.com](https://comfygomobility.com)**



## Assembling the Headrest

- Remove the padding from the top of the backrest.



- Position the headrest at the top of the backrest, ensuring the holes are aligned.
- Insert the two welded bolts into the headrest frame and secure the headrest to the backrest using the provided quick-release nuts.
- Reinstall the backrest padding for added comfort.



## Disassembling the Headrest

- Remove the padding from the top of the backrest.
- Release the two quick-release nuts and detach the headrest from the frame.
- Reinstall the backrest padding for added comfort.




## Adjusting the Height of the Headrest

- The headrest is height-adjustable.
- To adjust its height, hold the headrest with one hand and apply pressure as needed.



# BATTERY & CHARGER



## Installing the Battery

1. Place the battery back onto the frame.
2. Secure the battery using the battery knob.
3. Connect the battery plug to the controller plug.

## Dismantling the Battery

1. First, disconnect the battery plug and the controller plug.
2. Release the battery by using the battery knob.

## Controller Plug

Charging  
Port

Battery  
Knob

## Charger Requirements

- The charger should only be used for charging the batteries.
- Do not use your wheelchair while it is charging.

## Charger Technical Data

- Input voltage: AC220V (110V)  $\pm 10\%$
- Output voltage: 24V/DC
- Output current: 1.5~5A

## Charger Use and Care

- A full charge requires 6 to 8 hours.
- Do not overcharge for more than 12 hours.
- To prevent shortening your battery's life, charge at least once per month when not in use.
- Do not stop charging until the process is complete.
- Repeated use of less than fully charged batteries will shorten their lifespan.
- When fully charged, the power indicator turns green. Do not stop charging until this indicator is lit.
- Turn off the power supply after charging to prevent battery discharge.
- Avoid over-charging for more than 12 hours as it is dangerous.

## Follow These Rules to Avoid Issues with Chargers:

- Use a national standard charger with an output voltage of 24V/DC 1.5A~5A if the wheelchair does not include one.
- Ensure the charger is well ventilated during charging.
- Do not expose the wheelchair to direct sunlight or humidity.
- Charge within a temperature range of 50 to 122 °F (10 to 50 °C).



# BATTERY & CHARGER

- Never use or sit on the wheelchair while it is charging.
- It is normal for the charger's fan to make sound during charging.
- Prevent liquid from entering the charger.
- Do not place the charger on flammable items.
- Stay away from flames while charging.
- Do not smoke while charging.
- Avoid unplugging the charger with wet hands or from a wet socket.

## Charging Directly While in Battery

- Make sure the charger groove is not blocked.
- Ensure the wheelchair's power is off.
- Unplug connections between the battery box and the controller.
- Connect the charger's output plug to the battery box's power plug.
- Connect the charger plug to the power supply; the red light will light up.

## Charging Through Your Controller

- Ensure the charger groove is not blocked.
- Ensure the wheelchair is turned off.
- Make sure the battery and controller are connected.
- Connect the charger's output plug to the socket below the controller.
- Connect the charger's main plug to the power supply; the red light will be alight.

## Battery Use and Maintenance

- Incorrect insertion of the battery may cause an explosion risk. Use only the provided batteries.
- While inserting the battery, ensure correct pole placement. Always follow the charging instructions provided here for optimal battery life.
- For replacement batteries, visit **[www.ComfyGoMobility.com](http://www.ComfyGoMobility.com)** or contact us at **[support@ComfyGoMobility.com](mailto:support@ComfyGoMobility.com)**

## A Few Ways to Keep Your Batteries Healthy:

- Store your batteries at room temperature. Excessive heat is the primary factor in reducing battery life.
- Remember, batteries deteriorate over time, regardless of usage.
- Avoid completely discharging the batteries, as this can lead to performance deterioration.
- Gradual performance decline can occur if the battery is left in a discharged state for extended periods.





# FEATURES



## Flipping-Up Your Armrest

- To flip up the armrest, press the gray armrest button with one hand and use the other hand to lift the armrest.
- Press down on the armrest to return it to its proper position and secure it.

Armrest button



## Installing and Dismantling the Anti-tipper

- Pull the quick-release to install or dismantle the anti-tipper.
- When using the chair, adjust the anti-tipper to the longest position for your safety.

**!** *WARNING: Failing to properly assemble the anti-tip wheels is a safety violation. Ensure that the anti-tip wheels are correctly assembled.*



## Storage Compartments

- Your power chair comes equipped with two storage compartments: one located on the seatback and the other between the seat and the foot platform.

**!** *WARNING: Do not store personal property or valuables in any storage compartment when left unattended.*

**!** *WARNING: The storage compartment on the rear of the seatback is designed for small personal items. Avoid storing sharp objects in the compartments.*

**!** *WARNING! Do not carry heavy items weighing more than 5 lbs. (2.3 kg) in the storage compartments. Exceeding this weight limit may affect the chair's center of gravity, potentially causing a tip or fall.*






# FEATURES



## Footplate

- The footplate can be pulled up or down



## Seat Belt (Positioning Belt)

- Your power chair may come equipped with a seat belt, which can be adjusted for operator comfort. The seat belt is intended to support the operator to prevent sliding down or forward in the seat. It is not designed as a restraining device
- To Adjust the Seat Belt: Once seated, insert the metal tab on one side of the belt into the plastic housing on the opposite side until you hear a click. Pull the excess strap attached to the metal tab until secure, but ensure it's not too tight to cause discomfort.

- To Release the Positioning Belt: Press the push-button mechanism on the plastic housing.
- The seat belt (positioning belt) should always be secured. Never allow it hang on drag on the floor, as it may become entangled.

**!** *WARNING: The seat belt (positioning belt) should always be secured. Never allow it hang on drag on the floor, as it may become entangled.*

**!** *WARNING! The seat belt (positioning belt) is not intended for use as a seat belt in a motor vehicle. Furthermore, your power chair is not suitable for use as a seat in any vehicle. Anyone traveling in a vehicle should be properly belted into seats approved by the vehicle manufacturer.*

**!** *WARNING! Ensure the seat belt (positioning belt) is properly secured to the power chair and adjusted for operator comfort before each use.*

**!** *WARNING! Inspect the seat belt (positioning belt) for loose parts, damage (including tears, worn spots, bent hardware, damaged latch mechanisms), dirt, or debris before each use of the power chair. If any issues are found, contact your dealer.*





# EXTENSION KIT INSTALLATION



- **Wheelchair Seat Extension Kit Options:**
- **1.25" (3.2 cm) Extension Kit:**
  - IQ-8000: Extends from 17.5" (45 cm) to 18.75" (48.2 cm)
  - IQ-8000 Plus: Extends from 20.5" (52 cm) to 21.75" (55.7 cm)
- **0.625" (1.6 cm) Extension Kit:**
  - IQ-8000: Extends from 17.5" (45 cm) to 18.125" (46.6 cm)
  - IQ-8000 Plus: Extends from 20.5" (52 cm) to 21.125" (53.6 cm)

## Step-by-Step Installation:

**Important:** Repeat the installation steps below for both armrests.



### 1. Prepare for Installation:

- Park the wheelchair on a flat, stable surface.
- Turn off the wheelchair and ensure it is fully stationary.
- Remove any objects from the seat.



### 2. Remove Armrests:

- Loosen and remove the screws or bolts connecting the armrests to the seat frame using a hex key or appropriate screwdriver.
- Carefully detach and set aside the armrests.



### 3. Install Extension Kit:

- Choose the appropriate extension kit based on the desired width.
- Slide the extension bars provided in the kit into the existing armrest connection points on the wheelchair frame.
- Align the holes of the extension bars with those on the frame.





# EXTENSION KIT INSTALLATION



## 4. Secure Extension Bars:

- Insert the provided bolts, washers, and spacers into the aligned holes.
- Tighten securely using a wrench or hex key.
- Do not overtighten, as this may damage components.



## 5. Reattach Armrests:

- Place the armrests onto the newly installed extension bars.
- Secure the armrests using provided bolts and washers.
- Ensure they are tightly fastened and stable.



## 6. Final Check:

- Ensure all bolts and fasteners are securely tightened.
- Check stability by gently shaking the armrests.
- Verify the seat width now matches the desired extended width.

 *Priority each use inspect bolts and fasteners for tightness and wear.*

 *Do not exceed the recommended seat width limits provided by the kit.*

 *If instability or excessive play is detected, stop using the wheelchair immediately and consult technical support.*

 *Daily basis inspect and retighten bolts as needed.*

 *Clean extension bars and fittings regularly to prevent dirt build-up and ensure optimal performance.*

For technical assistance or additional support, please email us at **info@comfygomobility.com**, or call **+1 844 755 5055**.





# SYSTEM DIAGNOSTICS

The LED indicators on your device provide signals for various technical malfunctions which may affect parts such as the motor, brakes, battery, and wire connections. The table below will guide you through troubleshooting specific technical issues. While these issues can usually be detected without additional service tools, please do not hesitate to contact us if further assistance is required.

Refer to the table for interpretations of different audio signals.

Description of LED Light Indicators	Interpretation Guidelines	Issues and Solutions
All LED lights are off, and there is no sound.	The power has been turned off, or the wheelchair is in standby or sleep mode.	Power is not connected, or there may be a short fuse or a burnout.
Condition: All LED lights are lit.	The power is turned on and the system is performing self-diagnostics.	None
LED lights are partially off.	The lights indicate the remaining battery power.	Lights being partially off indicates that the battery power is only partially available.
One short beep and LED lights blinking once.	The battery power is very low.	The unit should be charged before use, or the battery may be malfunctioning and unable to take a charge
Two short beeps and LED lights blinking twice.	The left motor is malfunctioning.	<div><div>1.</div><div>The left motor connection might be broken or the wire could be disconnected.</div><div>2.</div><div>Please swap the wiring harnesses of the left and right motors. If you hear four alarms, there is a motor circuit fault. If there are two alarms, it indicates a controller circuit fault.</div></div>
Three short beeps and LED lights blinking three times.	The left magnetic brake is malfunctioning.	<div><div>1.</div><div>The left magnetic brake is in manual mode. Please switch to electric mode.</div><div>2.</div><div>Please check if the wiring harness of the magnetic brake is intact.</div><div>3.</div><div>Please swap the wiring harnesses of the left and right motors. If you hear five alarms, it indicates a left motor brake fault. If there are three alarms, it signifies a controller brake circuit fault.</div></div>
Four short beeps and LED lights blinking four times.	The right motor is malfunctioning.	<div><div>1.</div><div>The right motor connection might be broken or the wire could be disconnected.</div><div>2.</div><div>Please swap the wiring harnesses of the left and right motors. If you hear two alarms, there is a motor circuit fault. If there are four alarms, it indicates a controller circuit fault.</div></div>
Five short beeps and LED lights blinking five times.	The right magnetic brake is malfunctioning.	<div><div>1.</div><div>The right magnetic brake is in manual mode. Please switch to electric mode.</div><div>2.</div><div>Please check if the wiring harness of the magnetic brake is intact.</div><div>3.</div><div>Please swap the wiring harnesses of the left and right motors. If there are three alarms, it indicates a left motor brake fault. If there are five alarms, it signifies a controller brake circuit fault.</div></div>





# SYSTEM DIAGNOSTICS

Description of LED Light Indicators	Interpretation Guidelines	Issues and Solutions
Six short beeps and LED lights blinking six times.	The motor overcurrent	<ol style="list-style-type: none"><li>Motor overcurrent protection alarm.</li><li>When the alarm sounds, the gear indicator light showing 1st gear indicates an overcurrent in the left motor, and 5th gear indicates an overcurrent in the right motor.</li><li>Please disconnect the wiring of the left and right motors from the controller, turn on the power to the controller, and immediately push the joystick. If no alarm sounds, or if alarms 2, 3, 4, or 5 sound, it indicates a short circuit in the motor. If 6 alarms sound, it indicates a short circuit in the controller circuit.</li></ol>
Seven short beeps and LED lights blinking seven times.	The joystick is not in the initial position.	<ol style="list-style-type: none"><li>Turn off the power, return the joystick to the center position, and then turn the power back on.</li><li>If the joystick is in its initial position but emits 7 alarms, this indicates a controller circuit malfunction.</li></ol>
Eight short beeps and LED lights blinking eight times.	The controller system failure	Please refer to the Joystick (controller) failure table below for more information. Consult your dealer for maintenance.
Nine short beeps and LED lights blinking nine times.	The controller overloadt	Controller overload occurs when wheelchairs encounter excessively steep slopes or heavy loads while climbing slopes. Turning off the controller and then turning the power back on again can resolve the fault.
Ten short beeps and LED lights blinking ten times.	The battery failure	<ol style="list-style-type: none"><li>The battery power is very low. Please charge the battery.</li><li>Please check whether the plug connected to the battery is normal and whether the battery output voltage is normal.</li></ol>
Eleven short beeps and LED lights blinking Eleven times.	The controller is overheating	Turn off the controller and allow it to cool down for a period of time before using it again.












# SYSTEM DIAGNOSTICS

## Joystick (Controller) Failure Table

The 8-sound alarm (eight short beeps) indicates a controller system failure.  
The fault can be accurately located using the table provided..

Illustration	Interpretation	Issue
	Power butoon failure	Controller power issue.
	Parameter storage fail- ure.	Parameter storage exception detected.
	Serial port failure	Communication failure between boards. This indicates a con- troller system malfunction.
	Joystick Signal Failure	Joystick component circuit malfunction. Abnormal joystick signal caused by water ingress into the controller and external damage to the joystick components.
	Relay failure	Malfunctions caused by faulty relays.
	The deviation in the left current sampling is too large.	The left current sampling circuit is faulty.
	The deviation in the right current sampling is too large.	The right current sampling circuit is faulty.





# MAINTENANCE AREAS

## Maintaining Your Wheelchair

- Maintenance of your wheelchair should include cleaning, checking the wheels and battery, and charging the batteries.
- For detailed maintenance information, please contact your dealer.
- It is recommended that you have your wheelchair inspected every six months and schedule an annual service through your dealer.

## Cleaning Your Wheelchair

- Regularly clean your wheelchair, particularly the parts frequently touched by the user (such as the seat cushion, armrest, and controller), using a clean and slightly damp cloth.
- Please refrain from using organic solvents for cleaning, as these might damage the surface.
- The wheelchair should be cleaned approximately once per week.
- If used by an infectious patient, it should be cleaned and disinfected with suitable disinfectants after each use.

## Tires

- Regularly check the tire air pressure and the condition for wear.
- Tires require replacement when the tread pattern depth is reduced to 1mm.

## Batteries

- Ensure that the batteries are fully charged.
- To prolong battery life, it is strongly recommended to not charge the unit until the batteries are nearly depleted (when red or yellow LED indicators are lit).
- Also, refer to battery & charger section of this manual for detailed battery maintenance and charging instructions.

## Wheelchair Malfunction Signals

- If malfunctions occur during the operation of the wheelchair, turn off the power before inspecting the unit.

## Symptom

- Complete loss of power, and all the LED lights on the controller panel are off.

## Instructions:

- Step 1: Check if the controller plug is loose.
- Step 2: Check the connection between the controller plug and the battery box.





# MAINTENANCE AREAS

- Reinsert the plug by holding the plug itself when pulling it out. Do not pull on the wire to avoid damage.
- If power is not restored after these steps, or if you have any questions, please contact your dealer.
- The controller has a diagnostic system to monitor itself and the motor. The controller indicates any malfunctions of these parts. For more details, please refer to the chapter on audio signal indication (system diagnostics).

## Maintenance Checking

Your electric wheelchair should be checked every six months and serviced annually, as recommended.

Continuous attention is required for the commonly used elements, which are listed below:

### Area A

**Before each use, check if the following parts are functioning correctly:**

1. Backrest
2. Armrests
3. Controller Position
4. Footrest
5. Nuts and bolts
6. Battery power
7. Clutch/adjusting lever for conversion between manual and electric modes

### Area B

**The following parts should be checked monthly to prevent loosening or excessive wear:**

1. Screws
2. Brakes
3. Clutch/adjusting lever for conversion between manual and electric modes
4. The front and rear wheels, including tread
5. Connectors of the controller and charger





# WARRANTY POLICY

## Introduction

- We are dedicated to ensuring your satisfaction with our electric wheelchair. This warranty policy outlines the terms of coverage for your purchase.

## Coverage

- The frame of the electric wheelchair is covered by a three-year limited warranty.
- The motor, battery, and electronic components are covered by a one-year limited warranty.
- Spare parts, excluding those subject to normal wear and tear, are covered by a one-year warranty.

## Exclusions

- The following are not covered by the limited warranty:
  - Labor costs associated with part installation, and device failure due to misuse or negligence.
  - Non-durable components and parts that undergo normal wear and tear, including but not limited to tires, fuses, and lubricants.
  - Costs for routine service maintenance.
  - Unauthorized alterations or repairs to the wheelchair.
  - Damage from external sources or natural disasters.
  - Normal Wear and Tear
  - Parts that fail or wear due to normal use and aging, which require periodic replacement, are not covered under this warranty.

## Claim Process

- To file a warranty claim:
  - Contact our customer support with your proof of purchase and the wheelchair's serial number.
  - Follow the guidance provided by our support team regarding shipment and documentation for the parts in question.

## Parts Repair or Replacement

- During the warranty period, defective parts acknowledged by our technicians will be repaired or replaced. The owner will bear the labor costs for the installation of these parts. Replacement parts may be new or refurbished to the manufacturer's standards.

## Shipping Costs

- Shipping costs for warranty service are the responsibility of the wheelchair owner, except as required by local consumer protection laws.  
service team.

## Service Turnaround Time

- We aim to complete warranty services within a reasonable timeframe, typically within 30 days of





# WARRANTY POLICY

receiving the defective component.

## Warranty Transfer

- This warranty is non-transferable and applies only to the original purchaser of the wheelchair.

## Limitation of Damages

Our liability under this warranty is limited to the provision of replacement parts. We are not liable for incidental or consequential damages.

## Governing Law

This warranty shall be governed by and construed in accordance with the laws of the State of California, except where local laws mandate otherwise.

## Extended Protection Plans

We offer additional protection plans beyond the standard warranty coverage. For details, please contact our sales team.

## End of Life Handling

For environmental protection, do not dispose of the wheelchair or its batteries as household waste. Follow local disposal regulations, or return them to our service center.

## Effective Date

This warranty policy is effective as of 10/31/2023 and applies to all products purchased on or after this date.

## Changes to Policy

We reserve the right to change this warranty policy at any time. Customers will be informed of any changes via our website or direct communication.

## International Warranty

International warranty claims should be directed to the local dealer or our international customer service team.

## Contact Information

For questions or to initiate a warranty claim, please reach out to us at:

Phone: **+1 844 755 5055**

Email: **[info@ComfyGoMobility.com](mailto:info@ComfyGoMobility.com)**

Warranty Registration: **[Comfygomobility.com/pages/product-registration](https://Comfygomobility.com/pages/product-registration)**







*makes life easier*

**comfygomobility.com**

**United States of America & Canada**

- **[info@ComfyGoMobility.com](mailto:info@ComfyGoMobility.com)**
- **+1 844 755 5055**
- Visit **[comfygomobility.com/pages/contact](https://comfygomobility.com/pages/contact)** or scan the QR code with your smartphone to fill out the help form.



**Europe**

- **[info@ComfyGoMobility.eu](mailto:info@ComfyGoMobility.eu)**
- **+31 (0) 20 809 7031**
- Visit **[comfygomobility.eu/pages/contact](https://comfygomobility.eu/pages/contact)** or scan the QR code with your smartphone to fill out the help form.